

PHILIPS

PageWriter Trim

XML ECG Receiver

(available in English only)

Click on a blue button to select a file

This software is provided as-is, without any express or implied warranty. In no event shall Philips Medical Systems be held liable for any damages arising from the use of this software.

The XML ECG Receiver application is used to receive ECGs sent from a Philips PageWriter Trim or PageWriterTouch cardiograph. The transmission of ECGs from the cardiograph to the receiving computer is similar to the transmission of ECGs to a Philips TraceMaster or TraceMasterVue ECG Management System.

The XML ECG Receiver operates with Windows Information Services (IIS) as an IIS filter. The IIS filter detects ECG files sent by the http protocol, and stores the ECG files in the specified ECG directory on the receiving computer.

Please note that the networking configuration option on the cardiograph must be properly configured with the correct Computer Name or IP Address of the receiving computer.

The XML ECG Receiver application operates with Microsoft XML Core Services (MSXML 4.0 library) version 4.0 or higher. Click the button below for more information.

More Info — Click to read the program documentation as a Readme.txt file

More Info — Click to run the setup.exe installer

More Info — Click to read the program documentation for the Microsoft MSXML 4.0 Service Pack 2 program

---